

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2509

Name: RITTER2  
Sterilizer ID: V2313228  
06/13/2023 12:59 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.7 F 0.0  
2:00 194.3 F 4.3  
4:00 208.5 F 9.7  
6:00 225.7 F 6.7  
8:00 236.8 F 11.5  
10:00 234.9 F 9.4  
12:00 238.0 F 9.2  
14:00 249.1 F 15.5  
16:00 261.6 F 22.5  
18:00 272.5 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.3  
1:00 272.6 F 29.6  
2:00 272.0 F 29.8  
3:00 272.3 F 29.4  
4:00 272.1 F 29.9  
  
Min 271.6 F 29.1  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.6 F 0.0  
5:00 232.7 F 0.0  
10:00 239.1 F 0.0  
15:00 228.8 F 0.0  
20:00 223.1 F 0.0  
25:00 219.5 F 0.0  
30:00 225.3 F 0.0  
Drying Complete

FILLING: 2:12  
HEATING: 18:24  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:57:53  
CYCLE COMPLETE

06/13/2023 01:56 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_