
Midmark M11 - v1.0.0
Total Cycles: 5352

Name: RITTER 1
Sterilizer ID: V2178198
02/20/2025 06:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.3 F 0.0
2:00 178.0 F 3.1
4:00 199.1 F 7.7
6:00 219.6 F 13.2
8:00 227.0 F 6.9
10:00 237.5 F 11.5
12:00 233.6 F 8.1
14:00 243.9 F 12.1
16:00 244.5 F 12.4
18:00 256.4 F 18.8
20:00 267.7 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.4 F 29.0
2:00 271.6 F 28.3
3:00 271.8 F 28.9
4:00 272.1 F 28.9

Min 271.1 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.1 F 0.0
5:00 218.5 F 0.0
10:00 207.2 F 0.0
15:00 205.8 F 0.0
20:00 213.4 F 0.0
25:00 213.5 F 0.0
30:00 215.3 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 21:23
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:00:56
CYCLE COMPLETE

02/20/2025 07:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____