
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000501
11:37 07/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:35
STERILIZING 3:44
136.0 C 228kPa 3:44
136.2 C 232kPa 4:14
136.2 C 232kPa 4:44
136.4 C 233kPa 5:14
136.2 C 231kPa 5:44
136.4 C 233kPa 6:14
136.4 C 233kPa 6:44
136.3 C 232kPa 7:14
136.3 C 232kPa 7:14
VENTING 7:14
AIR DRYING 8:09
CYCLE COMPLETE 8:17

Digital Signature #
89D87AA45BFC12F5
