
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002408
7:18 12/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:44
PRESSURIZING 1:48
STERILIZING 2:42
136.0 C 220kPa 2:42
136.2 C 221kPa 3:12
136.4 C 223kPa 3:42
136.3 C 223kPa 4:12
135.9 C 221kPa 4:42
136.4 C 222kPa 5:12
136.4 C 224kPa 5:42
136.4 C 221kPa 6:12
136.4 C 221kPa 6:12
VENTING 6:12
AIR DRYING 7:05
CYCLE COMPLETE 7:14

Digital Signature #
7AA9B16A1F3F59B4
