
Midmark M11 - v1.0.0
Total Cycles: 5367

Name: RITTER 1
Sterilizer ID: V2178198
02/26/2025 11:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.4 F 0.0
2:00 171.6 F 2.4
4:00 170.9 F 5.8
6:00 203.8 F 10.0
8:00 223.9 F 5.8
10:00 219.2 F 8.2
12:00 229.9 F 12.7
14:00 233.8 F 9.0
16:00 240.9 F 11.1
18:00 244.0 F 12.3
20:00 255.1 F 18.1
22:00 265.3 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.2 F 29.1
2:00 271.4 F 28.9
3:00 272.0 F 29.1
4:00 271.2 F 28.9

Min 271.2 F 28.6
Max 272.2 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 196.5 F 0.0
10:00 202.3 F 0.0
15:00 201.8 F 0.0
20:00 200.9 F 0.0
25:00 205.1 F 0.0
30:00 210.7 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 23:56
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:03:28
CYCLE COMPLETE

02/26/2025 12:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____