
Midmark M11 - v1.0.5
Total Cycles: 761

Name: RITTER2
Sterilizer ID: V2313228
02/07/2022 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.2 F 0.0
2:00 98.2 F 0.0
4:00 116.0 F 0.1
6:00 154.6 F 1.6
8:00 181.1 F 5.2
10:00 204.0 F 9.3
12:00 216.6 F 9.5
14:00 225.3 F 8.7
16:00 236.0 F 12.0
18:00 239.5 F 11.4
20:00 238.3 F 9.9
22:00 251.8 F 16.6
24:00 263.4 F 23.4
26:00 272.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.7 F 29.6
2:00 271.8 F 28.9
3:00 271.7 F 29.3
4:00 272.5 F 29.5

Min 271.3 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.1 F 0.0
5:00 208.4 F 0.0
10:00 208.5 F 0.0
15:00 207.9 F 0.0
20:00 214.6 F 0.0
25:00 219.1 F 0.0
30:00 214.2 F 0.0

Drying Complete

FILLING: 0:54
HEATING: 26:12
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:04:11
CYCLE COMPLETE

02/07/2022 10:47 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____