

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2097

Name: RITTER2  
Sterilizer ID: V2313228  
02/27/2023 05:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.3 F 0.0  
2:00 154.1 F 3.0  
4:00 175.0 F 7.2  
6:00 203.9 F 12.8  
8:00 220.1 F 6.7  
10:00 234.6 F 12.2  
12:00 236.2 F 9.8  
14:00 236.4 F 8.8  
16:00 250.3 F 15.8  
18:00 262.9 F 23.0  
20:00 271.5 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.1 F 29.0  
2:00 272.6 F 29.5  
3:00 271.6 F 29.0  
4:00 272.4 F 29.2  
  
Min 271.5 F 28.7  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 170.9 F 0.0  
5:00 223.2 F 0.0  
10:00 230.9 F 0.0  
15:00 227.0 F 0.0  
20:00 225.8 F 0.0  
25:00 224.3 F 0.0  
30:00 228.5 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 20:01  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:58:35  
CYCLE COMPLETE

02/27/2023 06:44 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_