
Midmark M11 - v1.0.0
Total Cycles: 1769

Name: RITTER 1
Sterilizer ID: V2178198
03/30/2022 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.1 F 0.0
2:00 172.3 F 3.1
4:00 184.7 F 6.9
6:00 217.7 F 12.1
8:00 226.0 F 6.7
10:00 237.6 F 11.8
12:00 234.1 F 8.6
14:00 241.7 F 10.5
16:00 245.6 F 13.1
18:00 257.9 F 19.7
20:00 268.9 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 271.7 F 28.8
2:00 272.5 F 28.9
3:00 271.4 F 28.3
4:00 271.9 F 29.0

Min 271.1 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 237.1 F 0.0
10:00 229.6 F 0.0
15:00 217.8 F 0.0
20:00 210.9 F 0.0
25:00 214.0 F 0.0
30:00 217.5 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 21:01
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:20
CYCLE COMPLETE

03/30/2022 12:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____