
Midmark M11 - v1.0.5
Total Cycles: 1607

Name: RITTER2
Sterilizer ID: V2313228
10/17/2022 12:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.5 F 0.0
2:00 182.8 F 3.7
4:00 187.7 F 8.3
6:00 207.5 F 13.3
8:00 220.0 F 8.2
10:00 235.3 F 13.4
12:00 238.5 F 11.1
14:00 236.9 F 9.7
16:00 250.8 F 16.4
18:00 262.4 F 23.1
20:00 272.8 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.2
1:00 272.2 F 30.1
2:00 272.3 F 29.7
3:00 271.6 F 29.2
4:00 271.1 F 29.2

Min 271.0 F 29.0
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 221.2 F 0.0
10:00 227.3 F 0.0
15:00 231.1 F 0.0
20:00 227.4 F 0.0
25:00 226.3 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 20:27
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:23
CYCLE COMPLETE

10/17/2022 12:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____