
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 006603
17:33 13/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:45
STERILIZING 3:51
136.2 C 222kPa 3:51
136.3 C 224kPa 4:21
136.3 C 220kPa 4:51
136.0 C 223kPa 5:21
136.2 C 223kPa 5:51
136.4 C 224kPa 6:21
136.4 C 223kPa 6:51
136.4 C 223kPa 7:21
136.4 C 223kPa 7:21
VENTING 7:21
AIR DRYING 8:11
CYCLE COMPLETE 8:19

Digital Signature #
C558B7F2722B0ED4
