
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001451
10:49 12/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:42
STERILIZING 3:39
136.0 C 220kPa 3:39
136.2 C 223kPa 4:09
136.0 C 221kPa 4:39
136.4 C 222kPa 5:09
136.3 C 223kPa 5:39
136.4 C 225kPa 6:09
136.1 C 222kPa 6:39
136.3 C 224kPa 7:09
136.3 C 224kPa 7:09
VENTING 7:10
AIR DRYING 8:00
CYCLE COMPLETE 23:00

Digital Signature #
20B37F253C5402E4
