
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
2.2uS / 1.4ppm
CYCLE NUMBER 000018
16:00 13/01/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 4:02
STERILIZING 5:48
136.4 C 226kPa 5:48
136.4 C 228kPa 6:18
136.3 C 226kPa 6:48
136.4 C 227kPa 7:18
136.4 C 227kPa 7:48
136.4 C 226kPa 8:18
136.3 C 226kPa 8:49
136.4 C 227kPa 9:18
VENTING 9:18
AIR DRYING 10:36
CYCLE COMPLETE 12:36

Digital Signature #
FE6F1174A79F990A
