
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001449
10:08 01/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:50
STERILIZING 3:55
136.4 C 220kPa 3:55
136.3 C 223kPa 4:25
136.1 C 221kPa 4:55
135.9 C 221kPa 5:25
136.2 C 223kPa 5:55
136.3 C 222kPa 6:25
136.2 C 222kPa 6:55
136.4 C 225kPa 7:25
136.4 C 225kPa 7:25
VENTING 7:26
AIR DRYING 8:17
CYCLE COMPLETE 8:23

Digital Signature #
893122E346234BD1
