
Midmark M11 - v1.0.5
Total Cycles: 4510

Name: RITTER2
Sterilizer ID: V2313228
04/14/2025 01:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 81.6 F 0.0
2:00 82.1 F 0.3
4:00 84.3 F 1.6
6:00 133.6 F 4.1
8:00 154.3 F 7.2
10:00 170.5 F 10.8
12:00 203.6 F 5.0
14:00 215.2 F 6.1
16:00 223.7 F 9.8
18:00 234.7 F 10.3
20:00 236.6 F 10.3
22:00 236.9 F 9.2
24:00 246.8 F 14.3
26:00 257.7 F 20.3
28:00 267.8 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.6
1:00 272.4 F 29.8
2:00 272.1 F 30.3
3:00 271.3 F 29.5
4:00 272.3 F 30.1

Min 271.2 F 29.5
Max 272.6 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 205.5 F 0.0
10:00 217.5 F 0.0
15:00 186.9 F 0.0
20:00 181.0 F 0.0
25:00 177.3 F 0.0
30:00 170.9 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 29:20
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:08:07
CYCLE COMPLETE

04/14/2025 02:38 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____