

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4276

Name: RITTER 1  
Sterilizer ID: V2178198  
02/06/2024 02:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 102.0 F 0.0  
2:00 128.6 F 1.0  
4:00 165.3 F 3.1  
6:00 186.6 F 6.2  
8:00 202.6 F 9.7  
10:00 207.7 F 8.0  
12:00 220.7 F 5.2  
14:00 229.5 F 8.9  
16:00 238.8 F 13.1  
18:00 235.0 F 9.1  
20:00 242.8 F 11.1  
22:00 244.6 F 12.4  
24:00 256.2 F 18.5  
26:00 266.6 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.7 F 28.9  
2:00 271.7 F 28.1  
3:00 272.7 F 29.0  
4:00 272.1 F 28.4

Min 271.6 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.0 F 0.0  
5:00 229.4 F 0.0  
10:00 235.5 F 0.0  
15:00 227.5 F 0.0  
20:00 223.3 F 0.0  
25:00 224.8 F 0.0  
30:00 221.6 F 0.0

Drying Complete

FILLING: 0:57  
HEATING: 27:39  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:06:09  
CYCLE COMPLETE

02/06/2024 03:59 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_