
Midmark M11 - v1.0.0
Total Cycles: 2070

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2022 02:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.4 F 0.0
2:00 163.0 F 2.2
4:00 182.0 F 5.9
6:00 210.6 F 11.3
8:00 225.5 F 5.9
10:00 234.4 F 11.0
12:00 232.4 F 7.4
14:00 245.7 F 12.7
16:00 244.1 F 12.0
18:00 256.5 F 18.6
20:00 267.3 F 25.3
22:00 272.9 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.6
1:00 272.7 F 29.1
2:00 271.2 F 28.7
3:00 272.7 F 28.9
4:00 271.7 F 29.0

Min 271.2 F 28.5
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 234.9 F 0.0
10:00 207.9 F 0.0
15:00 222.1 F 0.0
20:00 213.8 F 0.0
25:00 234.2 F 0.0
30:00 232.1 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 22:06
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:49
CYCLE COMPLETE

06/28/2022 03:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____