
Midmark M11 - v1.0.0
Total Cycles: 3076

Name: RITTER 1
Sterilizer ID: V2178198
04/04/2023 03:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.4 F 0.0
2:00 186.6 F 3.6
4:00 206.4 F 8.3
6:00 225.0 F 11.9
8:00 231.3 F 9.3
10:00 236.4 F 9.2
12:00 243.1 F 12.6
14:00 242.2 F 11.4
16:00 255.5 F 18.2
18:00 267.0 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.7 F 29.0
2:00 271.7 F 28.2
3:00 272.0 F 28.9
4:00 272.3 F 28.8

Min 271.1 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.0 F 0.0
5:00 228.0 F 0.0
10:00 240.5 F 0.0
15:00 241.3 F 0.0
20:00 230.7 F 0.0
25:00 225.9 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

04/04/2023 04:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____