
Midmark M11 - v1.0.0
Total Cycles: 2546

Name: RITTER 1
Sterilizer ID: V2178198
11/11/2022 09:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.6 F 0.0
2:00 82.5 F 0.0
4:00 97.2 F 0.0
6:00 134.1 F 0.1
8:00 162.3 F 2.7
10:00 183.2 F 5.9
12:00 199.3 F 9.4
14:00 211.4 F 13.0
16:00 221.8 F 5.5
18:00 230.2 F 9.3
20:00 239.3 F 12.7
22:00 238.0 F 10.2
24:00 238.4 F 8.4
26:00 247.2 F 14.0
28:00 258.8 F 20.1
30:00 268.9 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.5 F 29.0
2:00 271.6 F 28.4
3:00 272.5 F 29.0
4:00 271.5 F 28.3

Min 271.5 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 214.7 F 0.0
10:00 208.9 F 0.0
15:00 218.0 F 0.0
20:00 216.3 F 0.0
25:00 221.0 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:12
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:12
CYCLE COMPLETE

11/11/2022 10:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____