
Midmark M11 - v1.0.0
Total Cycles: 3373

Name: RITTER 1
Sterilizer ID: V2178198
06/12/2023 11:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.1 F 0.0
2:00 179.2 F 4.0
4:00 190.7 F 8.3
6:00 209.8 F 13.5
8:00 223.7 F 7.6
10:00 237.1 F 13.0
12:00 237.7 F 10.3
14:00 235.2 F 8.4
16:00 250.2 F 15.5
18:00 262.6 F 22.4
20:00 272.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 271.8 F 28.8
2:00 272.6 F 29.0
3:00 271.9 F 28.5
4:00 271.2 F 28.2

Min 271.2 F 28.1
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 218.6 F 0.0
10:00 227.0 F 0.0
15:00 230.0 F 0.0
20:00 230.1 F 0.0
25:00 228.6 F 0.0
30:00 219.0 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 20:12
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:02
CYCLE COMPLETE

06/12/2023 12:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____