
Midmark M11 - v1.0.0
Total Cycles: 5429

Name: RITTER 1
Sterilizer ID: V2178198
03/21/2025 09:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.2 F 0.0
2:00 187.0 F 2.4
4:00 189.5 F 6.2
6:00 198.5 F 10.8
8:00 218.6 F 4.7
10:00 224.7 F 8.5
12:00 237.7 F 13.2
14:00 238.6 F 10.3
16:00 236.6 F 8.5
18:00 248.4 F 14.5
20:00 259.2 F 20.7
22:00 269.4 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 271.8 F 28.9
2:00 271.6 F 29.1
3:00 271.5 F 28.7
4:00 271.6 F 29.2

Min 271.0 F 28.5
Max 271.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 204.6 F 0.0
10:00 203.8 F 0.0
15:00 208.9 F 0.0
20:00 221.1 F 0.0
25:00 213.4 F 0.0
30:00 216.3 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 22:56
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:02:19
CYCLE COMPLETE

03/21/2025 10:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____