
Midmark M11 - v1.0.5
Total Cycles: 3143

Name: RITTER2
Sterilizer ID: V2313228
11/20/2023 11:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.3 F 0.0
2:00 172.5 F 1.3
4:00 186.7 F 5.0
6:00 211.8 F 9.8
8:00 223.0 F 8.2
10:00 231.3 F 9.5
12:00 238.5 F 9.7
14:00 242.4 F 12.6
16:00 242.5 F 11.7
18:00 255.1 F 18.3
20:00 266.4 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.5
2:00 271.8 F 28.8
3:00 271.7 F 29.3
4:00 272.5 F 29.5

Min 271.2 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 212.1 F 0.0
10:00 212.9 F 0.0
15:00 213.6 F 0.0
20:00 215.6 F 0.0
25:00 217.8 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 21:37
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:00:45
CYCLE COMPLETE

11/20/2023 12:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____