
Midmark M11 - v1.0.0
Total Cycles: 3088

Name: RITTER 1
Sterilizer ID: V2178198
04/11/2023 09:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.7 F 0.0
2:00 142.2 F 0.1
4:00 155.3 F 0.4
6:00 194.7 F 3.8
8:00 204.7 F 8.7
10:00 220.5 F 12.2
12:00 229.3 F 8.3
14:00 240.9 F 13.4
16:00 240.3 F 10.9
18:00 237.1 F 9.2
20:00 251.6 F 16.1
22:00 263.4 F 22.9
24:00 271.9 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.5 F 28.9
2:00 271.7 F 28.5
3:00 272.5 F 28.9
4:00 271.5 F 28.2

Min 271.5 F 28.2
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 219.9 F 0.0
10:00 245.4 F 0.0
15:00 240.3 F 0.0
20:00 229.4 F 0.0
25:00 228.7 F 0.0
30:00 230.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:06
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:02:11
CYCLE COMPLETE

04/11/2023 10:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____