
Midmark M11 - v1.0.0
Total Cycles: 2064

Name: RITTER 1
Sterilizer ID: V2178198
06/27/2022 04:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.6 F 0.0
2:00 182.7 F 3.0
4:00 194.7 F 7.3
6:00 213.6 F 12.7
8:00 225.5 F 6.4
10:00 238.9 F 11.7
12:00 235.9 F 9.3
14:00 237.8 F 9.1
16:00 249.3 F 15.0
18:00 262.0 F 22.2
20:00 272.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.1
2:00 271.6 F 28.3
3:00 272.6 F 29.3
4:00 272.0 F 28.7

Min 271.5 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 215.3 F 0.0
10:00 236.9 F 0.0
15:00 228.3 F 0.0
20:00 226.4 F 0.0
25:00 226.7 F 0.0
30:00 233.6 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 20:18
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:43
CYCLE COMPLETE

06/28/2022 06:55 AM
ACCEPTED: X
REJECTED:
Operator: U02