

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2800

Name: RITTER2  
Sterilizer ID: V2313228  
08/25/2023 04:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.6 F 0.0  
2:00 158.8 F 0.0  
4:00 186.8 F 3.6  
6:00 208.8 F 8.0  
8:00 224.0 F 13.2  
10:00 230.4 F 9.1  
12:00 237.9 F 10.2  
14:00 243.8 F 13.1  
16:00 243.5 F 12.4  
18:00 256.5 F 19.3  
20:00 268.1 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 271.8 F 28.9  
2:00 272.2 F 29.1  
3:00 272.5 F 29.5  
4:00 272.1 F 29.6

Min 271.6 F 28.8  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.3 F 0.0  
5:00 215.2 F 0.0  
10:00 214.7 F 0.0  
15:00 212.6 F 0.0  
20:00 209.8 F 0.0  
25:00 208.5 F 0.0  
30:00 209.4 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 21:11  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:59:31  
CYCLE COMPLETE

08/25/2023 05:16 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_