
Midmark M11 - v1.0.0
Total Cycles: 2348

Name: RITTER 1
Sterilizer ID: V2178198
09/21/2022 09:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.2 F 0.0
2:00 182.4 F 3.1
4:00 182.4 F 7.6
6:00 204.1 F 12.5
8:00 219.2 F 6.6
10:00 233.0 F 11.8
12:00 232.8 F 8.5
14:00 241.4 F 10.4
16:00 245.5 F 13.2
18:00 258.2 F 20.1
20:00 269.6 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.7 F 29.4
2:00 271.8 F 28.7
3:00 272.1 F 29.2
4:00 272.5 F 29.2

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.3 F 0.0
5:00 234.5 F 0.0
10:00 226.6 F 0.0
15:00 229.4 F 0.0
20:00 227.6 F 0.0
25:00 221.1 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 21:01
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:13
CYCLE COMPLETE

09/21/2022 09:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____