
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000029
11:20 09/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:44
STERILIZING 3:56
136.4 C 221kPa 3:56
136.4 C 226kPa 4:26
136.0 C 222kPa 4:56
136.4 C 221kPa 5:26
136.2 C 223kPa 5:56
135.8 C 220kPa 6:26
136.4 C 222kPa 6:56
136.4 C 221kPa 7:26
136.4 C 221kPa 7:26
VENTING 7:26
AIR DRYING 8:14
CYCLE COMPLETE 8:21

Digital Signature #
FF217ABD972D5933
