
Midmark M11 - v1.0.5
Total Cycles: 2936

Name: RITTER2
Sterilizer ID: V2313228
09/27/2023 10:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.6 F 0.0
2:00 171.8 F 2.1
4:00 183.7 F 6.4
6:00 204.7 F 11.8
8:00 222.7 F 7.3
10:00 238.8 F 13.0
12:00 238.3 F 10.5
14:00 236.1 F 8.9
16:00 251.0 F 15.9
18:00 263.1 F 23.0
20:00 271.3 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.9 F 29.6
2:00 271.6 F 29.3
3:00 273.0 F 29.6
4:00 271.4 F 29.1

Min 271.4 F 29.0
Max 273.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.4 F 0.0
5:00 235.5 F 0.0
10:00 202.8 F 0.0
15:00 183.6 F 0.0
20:00 180.1 F 0.0
25:00 177.8 F 0.0
30:00 178.0 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:52
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:39
CYCLE COMPLETE

09/27/2023 10:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____