
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007391
8:40 15/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:25
PRESSURIZING 3:26
STERILIZING 4:30
136.4 C 223kPa 4:30
135.8 C 220kPa 5:00
136.4 C 220kPa 5:30
136.3 C 221kPa 6:00
136.4 C 222kPa 6:30
136.4 C 224kPa 7:00
136.1 C 222kPa 7:30
136.1 C 221kPa 8:00
136.1 C 221kPa 8:00
VENTING 8:00
AIR DRYING 8:49
CYCLE COMPLETE 27:09

Digital Signature #
51344DD57E1ED73B
