
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002323
13:56 21/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:12
PRESSURIZING 2:19
STERILIZING 3:19
136.0 C 221kPa 3:19
136.2 C 223kPa 3:49
136.2 C 221kPa 4:19
136.2 C 224kPa 4:49
136.3 C 221kPa 5:19
136.4 C 224kPa 5:49
136.2 C 224kPa 6:19
136.2 C 223kPa 6:49
136.2 C 223kPa 6:49
VENTING 6:49
AIR DRYING 7:41
CYCLE COMPLETE 37:41

Digital Signature #
602DBD88D54551A3
