
Midmark M11 - v1.0.5
Total Cycles: 1239

Name: RITTER2
Sterilizer ID: V2313228
06/29/2022 08:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.1 F 0.0
2:00 158.5 F 1.9
4:00 160.0 F 5.1
6:00 180.6 F 9.5
8:00 206.1 F 8.0
10:00 220.6 F 8.5
12:00 233.9 F 12.1
14:00 235.1 F 10.8
16:00 235.6 F 8.9
18:00 248.6 F 15.4
20:00 260.3 F 21.9
22:00 270.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.7 F 30.2
2:00 272.2 F 29.7
3:00 271.2 F 29.3
4:00 272.1 F 30.1

Min 271.2 F 29.2
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 210.7 F 0.0
10:00 226.2 F 0.0
15:00 222.3 F 0.0
20:00 219.9 F 0.0
25:00 227.9 F 0.0
30:00 230.6 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 22:51
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:55
CYCLE COMPLETE

06/29/2022 09:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____