
Midmark M11 - v1.0.5
Total Cycles: 249

Name: RITTER2
Sterilizer ID: V2313228
07/27/2021 06:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.6 F 0.0
2:00 74.0 F 0.0
4:00 79.5 F 0.2
6:00 112.4 F 1.9
8:00 143.3 F 4.8
10:00 172.2 F 8.3
12:00 201.9 F 12.2
14:00 216.9 F 5.6
16:00 223.1 F 9.7
18:00 231.5 F 10.8
20:00 236.6 F 10.4
22:00 235.0 F 8.5
24:00 247.0 F 14.3
26:00 258.0 F 20.4
28:00 268.1 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.3 F 29.8
2:00 271.4 F 29.0
3:00 272.2 F 30.0
4:00 272.1 F 29.4

Min 271.2 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.0 F 0.0
5:00 219.9 F 0.0
10:00 209.1 F 0.0
15:00 222.7 F 0.0
20:00 219.9 F 0.0
25:00 229.2 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 29:21
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:07:18
CYCLE COMPLETE

07/27/2021 07:27 AM
ACCEPTED: X
REJECTED:
Operator: U01