
Midmark M11 - v1.0.0
Total Cycles: 1789

Name: RITTER 1
Sterilizer ID: V2178198
04/04/2022 10:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.7 F 0.0
2:00 156.1 F 3.3
4:00 174.8 F 7.3
6:00 204.5 F 11.9
8:00 219.3 F 5.8
10:00 234.2 F 10.8
12:00 231.4 F 7.0
14:00 244.7 F 13.2
16:00 242.9 F 11.7
18:00 255.9 F 18.4
20:00 267.1 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.3
1:00 272.2 F 29.1
2:00 272.1 F 28.5
3:00 271.6 F 28.7
4:00 272.2 F 29.0

Min 271.0 F 28.1
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 234.5 F 0.0
10:00 195.1 F 0.0
15:00 184.0 F 0.0
20:00 185.3 F 0.0
25:00 190.1 F 0.0
30:00 192.8 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 21:19
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

04/04/2022 12:34 PM
ACCEPTED: X
REJECTED:
Operator: U02