
Midmark M11 - v1.0.0
Total Cycles: 1061

Name: RITTER 1
Sterilizer ID: V2178198
07/27/2021 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.8 F 0.0
2:00 189.4 F 4.4
4:00 212.1 F 10.7
6:00 230.5 F 6.8
8:00 242.1 F 11.6
10:00 245.4 F 12.7
12:00 244.3 F 12.1
14:00 258.8 F 19.8
16:00 271.3 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.3 F 28.4
2:00 271.2 F 28.3
3:00 272.3 F 28.5
4:00 272.9 F 28.8

Min 271.0 F 27.8
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.4 F 0.0
5:00 235.5 F 0.0
10:00 226.3 F 0.0
15:00 222.0 F 0.0
20:00 218.0 F 0.0
25:00 216.4 F 0.0
30:00 186.9 F 0.0
Drying Complete

FILLING: 2:13
HEATING: 16:37
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 00:55:51
CYCLE COMPLETE

07/27/2021 02:09 PM
ACCEPTED: X
REJECTED:
Operator: U01