
Midmark M11 - v1.0.5
Total Cycles: 4261

Name: RITTER2
Sterilizer ID: V2313228
01/13/2025 02:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.5 F 0.0
2:00 160.8 F 0.1
4:00 175.6 F 1.0
6:00 198.9 F 5.0
8:00 213.6 F 9.6
10:00 226.8 F 8.9
12:00 233.8 F 10.3
14:00 234.5 F 8.4
16:00 243.2 F 12.0
18:00 245.2 F 13.3
20:00 257.6 F 19.9
22:00 268.7 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.5
2:00 271.5 F 29.0
3:00 272.8 F 29.5
4:00 271.5 F 28.8

Min 271.4 F 28.7
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.7 F 0.0
5:00 207.9 F 0.0
10:00 214.3 F 0.0
15:00 219.1 F 0.0
20:00 224.0 F 0.0
25:00 225.1 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 23:11
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:01:41
CYCLE COMPLETE

01/13/2025 03:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____