

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3580

Name: RITTER 1  
Sterilizer ID: V2178198  
07/25/2023 02:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.3 F 0.0  
2:00 165.2 F 0.9  
4:00 180.6 F 3.7  
6:00 209.3 F 7.1  
8:00 212.7 F 11.3  
10:00 219.8 F 4.7  
12:00 232.5 F 9.1  
14:00 241.4 F 13.5  
16:00 239.8 F 10.7  
18:00 237.4 F 8.4  
20:00 249.4 F 14.8  
22:00 260.0 F 20.7  
24:00 269.9 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.5 F 28.7  
2:00 271.4 F 28.1  
3:00 272.5 F 28.8  
4:00 271.7 F 28.2

Min 271.3 F 28.0  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.9 F 0.0  
5:00 212.8 F 0.0  
10:00 217.6 F 0.0  
15:00 219.8 F 0.0  
20:00 221.5 F 0.0  
25:00 223.4 F 0.0  
30:00 222.7 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 24:58  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:03:32  
CYCLE COMPLETE

07/25/2023 03:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_