
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004926
11:17 09/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:11
PRESSURIZING 1:16
STERILIZING 1:56
136.0 C 220kPa 1:56
136.4 C 221kPa 2:26
136.2 C 219kPa 2:56
136.4 C 223kPa 3:26
136.2 C 221kPa 3:56
136.3 C 224kPa 4:26
136.0 C 220kPa 4:56
136.3 C 223kPa 5:26
136.3 C 223kPa 5:26
VENTING 5:26
AIR DRYING 6:12
CYCLE COMPLETE 6:40

Digital Signature #
11345DBAC60ACD3A
