
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000275
9:26 23/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:49
PRESSURIZING 2:52
STERILIZING 3:52
136.0 C 220kPa 3:52
136.4 C 222kPa 4:22
136.0 C 221kPa 4:52
136.4 C 222kPa 5:22
136.3 C 223kPa 5:52
136.4 C 223kPa 6:22
136.4 C 223kPa 6:52
136.4 C 223kPa 7:22
136.4 C 223kPa 7:22
VENTING 7:23
AIR DRYING 8:11
CYCLE COMPLETE 9:13

Digital Signature #
AF7FBACB43E85D90
