
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008512
7:55 04/04/2025

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 1:59
STERILIZING 3:06
136.4 C 220kPa 3:06
136.4 C 223kPa 3:36
136.3 C 220kPa 4:06
136.4 C 222kPa 4:36
136.2 C 221kPa 5:06
136.3 C 222kPa 5:36
136.4 C 223kPa 6:06
136.2 C 221kPa 6:36
136.2 C 221kPa 6:36
VENTING 6:36
AIR DRYING 7:26
CYCLE COMPLETE 37:26

Digital Signature #
7084FB0111E894DD
