
Midmark M11 - v1.0.5
Total Cycles: 2419

Name: RITTER2
Sterilizer ID: V2313228
05/26/2023 09:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.3 F 0.0
2:00 80.0 F 0.0
4:00 112.6 F 0.0
6:00 121.8 F 0.0
8:00 159.2 F 0.7
10:00 179.7 F 4.0
12:00 200.0 F 7.6
14:00 214.7 F 11.4
16:00 224.3 F 5.2
18:00 232.8 F 9.7
20:00 239.4 F 10.5
22:00 240.3 F 11.5
24:00 238.1 F 9.7
26:00 251.0 F 16.0
28:00 262.0 F 22.4
30:00 272.1 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.8
1:00 272.2 F 29.6
2:00 271.7 F 28.7
3:00 272.5 F 29.6
4:00 271.9 F 28.9

Min 271.3 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 197.3 F 0.0
10:00 206.6 F 0.0
15:00 208.7 F 0.0
20:00 212.5 F 0.0
25:00 216.0 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:22
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:08:17
CYCLE COMPLETE

05/26/2023 10:10 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____