

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1971

Name: RITTER 1  
Sterilizer ID: V2178198  
05/30/2022 10:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.8 F 0.0  
2:00 194.1 F 4.2  
4:00 190.9 F 9.0  
6:00 211.5 F 9.6  
8:00 225.3 F 8.6  
10:00 239.3 F 11.5  
12:00 238.2 F 10.8  
14:00 236.7 F 8.7  
16:00 249.9 F 15.4  
18:00 261.8 F 22.1  
20:00 272.4 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.5 F 29.1  
2:00 271.4 F 28.4  
3:00 272.6 F 29.2  
4:00 271.8 F 28.6  
  
Min 271.3 F 28.4  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.9 F 0.0  
5:00 214.9 F 0.0  
10:00 223.8 F 0.0  
15:00 235.2 F 0.0  
20:00 234.1 F 0.0  
25:00 224.3 F 0.0  
30:00 212.7 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 20:28  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:51  
CYCLE COMPLETE

05/30/2022 11:11 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_