
Midmark M11 - v1.0.0
Total Cycles: 3423

Name: RITTER 1
Sterilizer ID: V2178198
06/21/2023 12:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.1 F 0.0
2:00 199.8 F 3.6
4:00 209.7 F 8.6
6:00 219.8 F 8.5
8:00 231.4 F 9.1
10:00 239.2 F 9.6
12:00 243.4 F 12.5
14:00 242.1 F 11.2
16:00 255.5 F 18.1
18:00 266.9 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.8 F 29.0
2:00 271.9 F 28.3
3:00 271.8 F 28.5
4:00 272.9 F 28.9

Min 271.5 F 28.0
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 225.1 F 0.0
10:00 227.7 F 0.0
15:00 233.0 F 0.0
20:00 234.2 F 0.0
25:00 224.4 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 19:32
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

06/21/2023 01:45 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____