

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000877  
10:30 20/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:54  
PRESSURIZING 3:01  
STERILIZING 4:12  
136.0 C 226kPa 4:12  
136.2 C 231kPa 4:42  
136.4 C 232kPa 5:12  
136.3 C 232kPa 5:42  
136.2 C 230kPa 6:12  
136.4 C 233kPa 6:42  
136.4 C 230kPa 7:12  
136.2 C 231kPa 7:42  
136.2 C 231kPa 7:42  
VENTING 7:42  
AIR DRYING 8:37  
CYCLE COMPLETE 13:48  
-----

Digital Signature #  
C62C8EF090FE77C9  
-----