
Midmark M11 - v1.0.5
Total Cycles: 3557

Name: RITTER2
Sterilizer ID: V2313228
04/24/2024 05:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.4 F 0.0
2:00 181.8 F 3.2
4:00 203.9 F 9.1
6:00 220.1 F 6.7
8:00 234.7 F 11.0
10:00 235.6 F 9.0
12:00 238.0 F 9.3
14:00 250.6 F 15.8
16:00 263.9 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.7 F 29.1
2:00 272.0 F 29.0
3:00 272.5 F 29.0
4:00 272.0 F 29.0

Min 271.6 F 28.4
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.1 F 0.0
5:00 235.7 F 0.0
10:00 224.4 F 0.0
15:00 215.8 F 0.0
20:00 208.8 F 0.0
25:00 199.9 F 0.0
30:00 201.0 F 0.0

Drying Complete

FILLING: 2:45
HEATING: 17:52
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:57:40
CYCLE COMPLETE

04/24/2024 06:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____