
Midmark M11 - v1.0.5
Total Cycles: 2947

Name: RITTER2
Sterilizer ID: V2313228
09/28/2023 03:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.0 F 0.0
2:00 190.0 F 3.6
4:00 202.5 F 9.1
6:00 222.4 F 7.7
8:00 231.4 F 10.7
10:00 233.4 F 8.4
12:00 241.0 F 10.6
14:00 247.5 F 14.2
16:00 259.8 F 21.1
18:00 270.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.1 F 29.0
2:00 272.5 F 29.5
3:00 271.6 F 28.7
4:00 272.4 F 29.1

Min 271.5 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 219.0 F 0.0
10:00 217.7 F 0.0
15:00 215.3 F 0.0
20:00 216.4 F 0.0
25:00 224.0 F 0.0
30:00 219.9 F 0.0
Drying Complete

FILLING: 2:31
HEATING: 18:41
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

09/28/2023 04:13 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____