
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002619
8:27 08/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:44
PRESSURIZING 1:48
STERILIZING 2:41
136.0 C 220kPa 2:41
136.4 C 222kPa 3:11
136.4 C 220kPa 3:41
136.4 C 223kPa 4:11
136.2 C 220kPa 4:41
136.4 C 223kPa 5:11
136.4 C 222kPa 5:41
136.1 C 221kPa 6:11
136.1 C 221kPa 6:11
VENTING 6:11
AIR DRYING 7:04
CYCLE COMPLETE 8:31

Digital Signature #
5CCDDD8395C5955D
