

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3805

Name: RITTER2  
Sterilizer ID: V2313228  
07/22/2024 12:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.3 F 0.0  
2:00 141.9 F 1.6  
4:00 155.4 F 4.4  
6:00 176.8 F 7.9  
8:00 196.5 F 11.8  
10:00 213.0 F 5.0  
12:00 221.7 F 9.1  
14:00 232.1 F 13.3  
16:00 233.2 F 9.6  
18:00 237.6 F 10.1  
20:00 244.8 F 13.2  
22:00 255.7 F 19.0  
24:00 265.8 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.1  
1:00 273.3 F 29.8  
2:00 272.6 F 30.0  
3:00 271.4 F 29.5  
4:00 272.3 F 29.6

Min 271.0 F 29.0  
Max 273.3 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.0 F 0.0  
5:00 194.3 F 0.0  
10:00 192.1 F 0.0  
15:00 188.2 F 0.0  
20:00 183.6 F 0.0  
25:00 179.8 F 0.0  
30:00 177.6 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 25:50  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:04:40  
CYCLE COMPLETE

07/22/2024 01:22 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_