

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000369  
9:04 21/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:27  
PRESSURIZING 2:34  
STERILIZING 3:45  
136.4 C 232kPa 3:45  
136.4 C 233kPa 4:15  
136.2 C 232kPa 4:45  
136.4 C 230kPa 5:15  
136.3 C 232kPa 5:45  
136.2 C 229kPa 6:15  
136.2 C 232kPa 6:45  
136.3 C 233kPa 7:15  
136.3 C 233kPa 7:15  
VENTING 7:16  
AIR DRYING 8:06  
CYCLE COMPLETE 8:28

-----  
Digital Signature #  
DF47A61F604866AD  
-----