
Midmark M11 - v1.0.5
Total Cycles: 667

Name: RITTER2
Sterilizer ID: V2313228
12/29/2021 03:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.3 F 0.0
2:00 191.9 F 4.1
4:00 199.6 F 9.7
6:00 221.4 F 6.8
8:00 234.0 F 11.7
10:00 236.1 F 10.0
12:00 235.9 F 9.1
14:00 251.3 F 16.5
16:00 263.7 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.2 F 29.6
2:00 271.3 F 29.2
3:00 272.2 F 29.6
4:00 271.2 F 28.9

Min 271.1 F 28.9
Max 272.4 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 217.0 F 0.0
10:00 236.3 F 0.0
15:00 236.7 F 0.0
20:00 230.0 F 0.0
25:00 222.6 F 0.0
30:00 219.1 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 17:58
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:22
CYCLE COMPLETE

12/30/2021 07:44 AM
ACCEPTED: X
REJECTED:
Operator: U02