
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000682
7:14 15/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:16
PRESSURIZING 1:23
STERILIZING 2:36
136.4 C 232kPa 2:36
136.4 C 232kPa 3:06
136.1 C 229kPa 3:36
136.4 C 229kPa 4:06
136.2 C 230kPa 4:36
136.3 C 230kPa 5:06
136.3 C 231kPa 5:36
136.2 C 230kPa 6:06
136.2 C 230kPa 6:06
VENTING 6:06
AIR DRYING 6:56
CYCLE COMPLETE 6:57

Digital Signature #
7797208F6B6770DC
