
Midmark M11 - v1.0.0
Total Cycles: 705

Name: MIDMARK
Sterilizer ID: V2178198
02/25/2021 04:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.9 F 0.0
2:00 160.1 F 2.3
4:00 171.6 F 6.2
6:00 198.6 F 11.6
8:00 216.8 F 6.6
10:00 230.1 F 12.4
12:00 234.9 F 9.4
14:00 237.0 F 8.6
16:00 248.3 F 14.4
18:00 261.0 F 21.5
20:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.9 F 29.2
2:00 272.4 F 28.9
3:00 272.2 F 28.8
4:00 272.2 F 28.8

Min 270.9 F 28.2
Max 273.1 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.5 F 0.0
5:00 228.5 F 0.0
10:00 222.7 F 0.0
15:00 225.7 F 0.0
20:00 231.5 F 0.0
25:00 230.4 F 0.0
Dry Cycle Aborted

FILLING: 1:15
HEATING: 20:24
STERILIZING: 4:00
VENTING: 2:55
DRYING: 29:01
TOTAL CYCLE: 00:57:35

02/25/2021 05:22 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____